

QQHM2.E210326 - POWER SUPPLIES, MEDICAL AND DENTAL - COMPONENT

Power Supplies, Medical and Dental - Component

See General Information for Power Supplies, Medical and Dental - Component

MDEXX GMBH

Zeppelinstrasse 30
28844 Weyhe, GERMANY

E210326

Model No.	Rated Input			Max Output							
	Volts	Hz	SC	V	A	VA	OC	SP	EP	FC	GC
4AM9991-0AA20-1N, TAM9991-0AA20-1N-Single Phase Transformer											
	-	-	-	-	-	-	-	-	-	-	-
4AP2596-0FH28-2EA0, TAP2596-0FH28-2EA0 Three-phase transformer											
	380-480ac	50-60	-	400	-	1.6k	-	60601-1	-	-	-
4AP2796-OEM30-3DM4, TAP2796-OEM30-3DM4 Three-phase transformer											
	480-380ac	50	-	412	3.6	-	-	60601-1	-	-	-
4AU4995 0AB28 8Y, TAU4995 0AB28 8Y Three-Phase, Safety Isolating Transformer											
	-	-	-	-	-	-	-	-	-	-	-
4AU4995-0AA78-8Y, TAU4995-0AA78-8Y Three-phase transformer											
	480-380ac	50/60	-	400	28.7	-	-	2601-1	32B	0	0
				157	20.5	-	-				
4AV9926-1CB00, TAV9926-1CB00 Three-phase autotransformer											
	480-400ac	50/60	-	400	14.45	10000	-	60601-1	-	0	1
4BU4595-0BD32-8A, TBU4595-0BD32-8A Three-phase autotransformer											
	200-362ac	50/60	-	400	63.7	44000	-	60601-1	-	0	0
				410	62.1	44000	-				
4BU4595-0BD42-8A, TBU4595-0BD42-8A Three-phase[*r]											
	200-382ac	50-60	-	400ac	63.7	44000	-	60601-1	-	-	-
				410ac	62.1	44000	-				
4FC9283-0LA00-1E, 4FC9283-0LA00-1F (Module 1), 4AY1978 (Module 2)											
	230ac	50/60	0	5	2	156	9	60601-1	-	0	0
				24	2	-	9				
4FC9283-0LA01-1E											
	230ac	50/60	0	10.4	2	453	9	60601-1	-	0	0
				30.8	2	-	9				
PDU, Power distribution unit											
	-	-	-	-	-	-	-	-	-	-	-

[*r] - Output values are rated.

Marking: Company name and model designation.

Last Updated on 2011-04-29

The appearance of a company's name or product in this database does not in itself assure that products so identified have been manufactured under UL's Follow-Up Service. Only those products bearing the UL Mark should be considered to be Certified and covered under UL's Follow-Up Service. Always look for the Mark on the product.

UL permits the reproduction of the material contained in the Online Certification Directory subject to the following conditions: 1. The Guide Information, Assemblies, Constructions, Designs, Systems, and/or Certifications (files) must be presented in their entirety and in a non-misleading manner, without any manipulation of the data (or drawings). 2. The statement "Reprinted from **Cookies on UL** Online Certification Directory with permission from UL" must appear adjacent to the extracted material. In addition, the reprinted material must include a copyright notice in the following format: "© 2019 UL LLC"

We use cookies to personalize content and ads, to provide social media features and to analyze our traffic. We also share information about your use of our site with our social media, advertising and analytics partners..
Learn More

> [Cookie Settings](#)

✓ [Accept](#)