

# QQHM8.E210326 - POWER SUPPLIES, MEDICAL AND DENTAL CERTIFIED FOR CANADA - COMPONENT

## Power Supplies, Medical and Dental Certified for Canada - Component

See General Information for Power Supplies, Medical and Dental Certified for Canada - Component

### MDEXX GMBH

Zeppelinstrasse 30  
28844 Weyhe, GERMANY

E210326

Model No.	Rated Input			Max Output							
	Volts	Hz	SC	V	A	VA	OC	SP	EP	FC	GC
<b>4AM9991-0AA20-1N, TAM9991-0AA20-1N-Single Phase Transformer</b>											
	-	-	-	-	-	-	-	-	-	-	-
<b>4AP2596-0FH28-2EA0, TAP2596-0FH28-2EA0 Three-phase transformer</b>											
	380-480ac	50-60	-	400	-	1.6k	-	60601-1	-	-	-
<b>4AP2796-OEM30-3DM4, TAP2796-OEM30-3DM4 Three-phase transformer</b>											
	480-380ac	50	-	412	3.6	-	-	60601-1	-	-	-
<b>4AU4995 0AB28 8Y, TAU4995 0AB28 8Y Three-Phase, Safety Isolating Transformer</b>											
	-	-	-	-	-	-	-	-	-	-	-
<b>4AU4995-0AA78-8Y, TAU4995-0AA78-8Y Three-phase transformer</b>											
	480-380ac	50/60	-	400	28.7	-	-	2601-1	32B	0	0
				157	20.5	-	-				
<b>4AV9926-1CB00, TAV9926-1CB00 Three-phase autotransformer</b>											
	480-400ac	50/60	-	400	14.45	10000	-	60601-1	-	0	1
<b>4BU4595-0BD32-8A, TBU4595-0BD32-8A Three-phase autotransformer</b>											
	200-362ac	50/60	-	400	63.7	44000	-	60601-1	-	0	0
				410	62.1	44000	-				
<b>4BU4595-0BD42-8A, TBU4595-0BD42-8A Three-phase[*r]</b>											
	200-382ac	50-60	-	400ac	63.7	44000	-	60601-1	-	-	-
				410ac	62.1	44000	-				
<b>4FC9283-0LA00-1E, 4FC9283-0LA00-1F (Module 1), 4AY1978 (Module 2)</b>											
	230ac	50/60	0	5	2	156	9	60601-1	-	0	0
				24	2	-	9				
<b>4FC9283-0LA01-1E</b>											
	230ac	50/60	0	10.4	2	453	9	60601-1	-	0	0
				30.8	2	-	9				
<b>PDU, Power distribution unit</b>											
	-	-	-	-	-	-	-	-	-	-	-

[\*r] - Output values are rated.

Marking: Company name, model designation and Recognized Component Mark for Canada,



Last Updated on 2011-04-29

The appearance of a company's name or product in this database does not in itself assure that products so identified have been manufactured under UL's Follow-Up Service. Only those products bearing the UL Mark should be considered to be Certified and covered under UL's Follow-Up Service. Always look for the Mark on the product.

UL permits the reproduction of the material contained in the Online Certification Directory subject to the following conditions: 1. The Guide Information, Assemblies, Constructions, Designs, System Applications (files) must be presented in their entirety and in a non-misleading manner, without any manipulation of the data (or drawings). 2. The statement "Reprinted from the Online Certifications Directory with permission from UL" must appear adjacent to the extracted material. In addition, the reprinted material must include a copyright notice in the following form: "Reprinted from the Online Certification Directory with permission from UL".

We use cookies to personalize content and ads, to provide social media features and to analyze our traffic. We also share information about your use of our site with our social media, advertising and analytics partners..  
Learn More

> [Cookie Settings](#)

[Accept](#)